

The Hellenic Institute of Metrology (EIM)
certifies that

IPROEB S.A. – Metrology Laboratory

Bucuresti, Str. Candiano Popescu, nr.1 sector 4 Romania, phone: +40 263 238 165
participated in the bilateral inter-laboratory comparison

EIM-TEM-76:

**« Comparison Calibration of a temperature chamber, according to
DKD R 5-7 (Method A)»**

that was organized by NQIS/EIM, and concluded in May 2023 with the results presented as follows:

Point of Measurement	Value of E_n ratio
50 °C	0,03

Remarks:

- The E_n ratio expresses the normalized deviation of the measurement result of the participating laboratory from the respective reference value, based on the uncertainties.*
- The performance of the laboratory is successful if the value of the E_n ratio is lower or equal to one.*



Thessaloniki, 31/05/2023

Evmorfia Kokkini
Head of Temperature - Humidity Laboratory

Άκυρο χωρίς σφραγίδα και ψηφιακή υπογραφή. / Invalid without seal and digital signature.

Επαλήθευση ψηφιακής υπογραφής με: / Verification of digital signature with: Adobe Acrobat Reader